

CERTIFICATE

Survey of 01 December 2023

You have fulfilled the requirements of the External Quality Assessment with the following analysis

HLA Diagnostics 07 - HLA-DQ (772):

Validity 12 months:

HLA DQA1*05

HLA DQB1*02

HLA DQB1*03:02

(R) analysis is subject to the RiliBÄK

Customer:

10074873

Clinomics Europe Kft.

Molekuláris Genetikai Laboratórium

Dr. Orsolya Bíró

Miháلكovics u. 10.

1094 Budapest

Düsseldorf, 11 December 2023



Prof. Dr. med. Michael Spannagl
(Head of Reference Institution)



Dr.-Ing. Nils Lachmann
(Adviser)

CERTIFICATE OF PARTICIPATION

Survey of 01 December 2023

You have participated in the External Quality Assessment with the following analysis

HLA Diagnostics 07 - HLA-DQ (772):

HLA DQA1*05
HLA DQB1*02
HLA DQB1*03:02

(R) analysis is subject to the RiliBÄK

Customer:
10074873
Clinomics Europe Kft.
Molekuláris Genetikai Laboratórium
Dr. Orsolya Bíró
Miháلكovics u. 10.
1094 Budapest

Düsseldorf, 11 December 2023



Prof. Dr. med. Michael Spannagl
(Head of Reference Institution)



Dr.-Ing. Nils Lachmann
(Adviser)

Listing and evaluation of the results

10074873: Clinomics Europe Kft.
Molekuláris Genetikai Laboratórium
Dr. Orsolya Biró

Survey of 01 December 2023

Adviser:
Dr.-Ing. Nils Lachmann

Charité-Universitätsmedizin Berlin, ZTB-Zentrum
für Transfusionsmedizin
und Zelltherapie Berlin gGmbH, Gewebetypisierung
am Campus Virchow-Klinikum
Augustenburger Platz 1
13353 Berlin

Tel.: +49 30 450 665 098

Fax: +49 30 450 553 918

Mail: nils.lachmann@charite.de

772

HLA Diagnostics 07 - HLA-DQ

Analyte	Sam- ple	Method	Manu- facturer	Device	Your specification(s)	Correct specification(s)	TV- Type	Meets criteria
HLA DQA1*05	61	388	ZY		negative	negative	M	+
	62				positive	positive	M	+
HLA DQB1*02	61	388	ZY		positive	positive	M	+
	62				positive	positive	M	+
HLA DQB1*03:02	61	388	ZY		negative	negative	M	+
	62				negative	negative	M	+

** BRAVO **



Individual summary of results

10074873: Clinomics Europe Kft.
Molekuláris Genetikai Laboratórium
Dr. Orsolya Biró

Survey of 01 December 2023

772

HLA Diagnostics 07 - HLA-DQ

HLA DQA1*05 (N = 64, Rate of success: 100%)

Sample 61

Collective	negative	total
all methods	64 ●	64

Sample 62

Collective	positive	total
all methods	64 ●	64

The point corresponds to the correct result, the horizontal bar corresponds to your specification, the vertical bar corresponds to your collective

HLA DQB1*02 (N = 64, Rate of success: 98,4%)

Sample 61

Collective	negative	positive	total
all methods	1	63 ●	64

Sample 62

Collective	negative	positive	total
all methods	1	63 ●	64

The point corresponds to the correct result, the horizontal bar corresponds to your specification, the vertical bar corresponds to your collective

HLA DQB1*03:02 (N = 64, Rate of success: 100%)

Sample 61

Collective	negative	total
all methods	64 ●	64

Sample 62

Collective	negative	total
all methods	64 ●	64

The point corresponds to the correct result, the horizontal bar corresponds to your specification, the vertical bar corresponds to your collective